# Sinuscan TM 301 ULTRASOUND FOR THE DIAGNOSIS OF SINUS DISEASE



# **SINUSCAN 301**

Sinuscan 301 is designed for detecting anomalies as for example fluid in the maxillary and frontal sinuses. This is done by the Sinuscan by indicating the back wall echo received from the bony back wall of a fluid filled cavity; no such echo is received if the cavity contains only air.

# PRINCIPLE OF OPERATION

The Sinuscan 301 works by transmitting ultrasonic energy pulses and receiving reflected sonar echoes from acoustic layers in the same manner as an echo sounder. The reflected sound wave is transformed into an electric signal and the signal is indicated by a graph on the OLED display as an A-mode curve.

The A-mode curve display indicates the layer distances and the strength of the echo. The resolution is 0.5 cm and the exploration depth is 7 cm. The ultrasonic frequency used (3 MHz) is transmitted through human soft tissue and bone, but not through air.

Svensk distributör och servicecenter:

**Entomed MedTech AB** 

Storgatan 7 233 31 Svedala 040-48 11 40

info@entomedmedtech.se www.entomedmedtech.se

# **FEATURES**

- Rapid examination, no need for time consuming preparations
- Safe for both the patient and the physician and painless, easy examination
- Also applicable to children from the age of three as well as to pregnant women
- The examination can easily be repeated with no additional risk for the patient
- According to scientific studies the method is very reliable
- Both left and right hand use can be applied
- 3 different exploration areas, Sin+Dex, Sin+Dex+Fro or Easy all-in-one mode
- Printing by USB-cable using Macros

# **TECHNICAL SPECIFICATIONS**

• Size: 210 x 69 x 36 mm

• Weight: 292 g with battery

• Battery: 6 V, 730 mAh, NiMH

• Display: OLED 30 x 57 mm

• Sensor: Ultrasound crystal 3 MHz, Ø 8 mm

Printing: by USB-cable using Macros

Sinuscan is a Trademark of Mediq Suomi Oy.



# **MANUFACTURER:**

Mediq Suomi Oy, P.O.Box 115, FI-O2201 Espoo, FINLAND, tel. +358 20 112 1744 fax +358 20 112 1501, e-mail: antti.tapanila@mediq.com, www.mediq.fi/sinuscan